

Name: Link Transportation Testing Technology (Shanghai) Company Limited

Address: Building 2, No. 778, Zhaoxian Road, Jiading District, Shanghai, China

Registration No. CNAS L6622

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS


Effective Date: 2023-04-18 Expiry Date: 2028-12-30

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1	*Dynamometer	Torque	《Torque Calibration》 WI-4.11-031, 《Pressure Calibration》 WI-4.11-032, 《Rotational Speed Calibration》 WI-4.11-033, 《Temperature Calibration》 WI-4.11-034, 《Volume Calibration》 WI-4.11-035, 《Length Calibration》 WI-4.11-036, 《Force Calibration》 WI-4.11-037, 《Air Speed Calibration》 WI-4.11-039, 《Humidity	(67~500) N·m	U= (0.4~2.9) N·m		
			(500~5640) N·m	U= (6.5~27) N·m			
			(5640~33900) N·m	U= (11.5~38.3) N·m			
		Pressure	(0.10~1.3) MPa	U= (0.001~0.004) MPa			
			(1.3~20) MPa	U= (0.003~0.037) MPa			
		Rotational speed	(20~51) MPa	U= (0.04~0.12) MPa			
		Temperature	(20 ~ 20000) r/min	U=(0.004~4)r/min			
		Volume	(-40 ~ 1300) °C	U=0.2 °C			
	(0.5~25) mL	U= (0.23~0.25) mL					



在线扫码获取验证

No. CNAS L6622

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Distance	Calibration》 WI-4.11-040	(0.01~2.5) mm	U=0.003mm		
				(25~150) mm	U=0.016mm		
		Force		(440~4440) N	U= (2.3~9.9) N		
				(4440~22200) N	U= (10.2~37.4) N		
				(22.2~155) kN	U= (0.07~0.47) kN		
		Air speed		(9.1~64) km/h	U= (1.8~2.3) km/h		
		Humidity		(10~95) %RH	U= (1.8~2.3) %RH		



No. CNAS L6622

The scope of the accreditation in Chinese remains the definitive version.