



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Link Engineering Company
43855 Plymouth Oaks Blvd.
Plymouth, MI 48170

Fulfills the requirements of

ISO/IEC 17025:2017

In the fields of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 21 October 2025

Certificate Number: ACT-1997.04



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Link Engineering Company

43855 Plymouth Oaks Blvd.
Plymouth, MI 48170

Daniel Williams d.williams@linkeng.com

CALIBRATION

Valid to: **October 21, 2025**

Certificate Number: **ACT-1997.04**

Electrical-DC and Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ⁴	Reference Standard, Method, and/or Equipment
DC Volt-Sourcing	Up to ±10 V	0.000 5 V	Direct Measurement using 6.4-L2-207 VMAX DAS
AC Volt-Sourcing	Up to 20 kHz Up to 10 Vpp	0.005 V	
Thermocouple-Sourcing	Types K, J, T (-40 to 2400) °F	0.86 °F	Electrical Simulation using 6.4-L2-207 VMAX DAS


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ⁴	Reference Standard, Method, and/or Equipment
Frequency-Measuring Equipment	Up to 20 kHz	0.57 Hz	Direct Measurement using 6.4-L2-207 VMAX DAS Calibration Setup

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is part of and must be included with the Certificate of Accreditation No. ACT-1997.04



Jason Stine, Vice President